

Certificate

Product Name : **Switching power supply**
Brand Name : **MEAN WELL**
Model Number : **NPF-120x-yz(x= blank or D ; y= 12, 15, 20, 24, 30, 36, 42, 48, 54 ; z= blank or BE),
PWM-120-xy(x= 12,24,36,48)**
Applicant : **MEAN WELL ENTERPRISES CO., LTD.**
Address : **No.28, Wuquan 3rd Rd., Wugu Dist., New Taipei City 248,Taiwan
(R.O.C.)**
Report Number : **S3C15-M020-1409-189**
Issue Date : **February 11, 2026**
Applicable Standards : **EN IEC 55015:2019/A11:2020
EN IEC 61000-3-2:2019
EN 61000-3-3:2013/A1:2019
EN IEC 61547:2023
EN 61000-4-2:2009
EN 61000-4-3:2006/A2:2010
EN 61000-4-4:2012
EN 61000-4-5:2014/A1:2017
EN 61000-4-6:2014
EN 61000-4-8:2010
EN IEC 61000-4-11:2020**



Based on the EMC Directive 2014/30/EU and the specifications of the customer, one sample of the designated product has been tested in our laboratory and found to be in compliance with the EMC standards cited above.



Central Research Technology Co.

11, Lane 41, Fushuen St., Jungshan Chiu,
Taipei 104, Taiwan
Tel : 886-2-25872719
Fax : 886-2-25850364

(Jim Chien/Authorized Signatory)

Date: February 11, 2026

TAF 0905
FCC CAB Code TW1104, TW0019
NVLAP Lab Code 200575-0

ISED CAB Code TW0905
VCCI Accep. No. R-11527, C-11609, T-11441, G-10010, C-20010
T-20009, G-10614