

Certificate



Certificate no.

TA 50220730 01

License Holder:

Mean Well Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

Mean Well Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Test report no.: ZTW1-BHU 10034233 001

Client Reference: 8208400364/JH110811

Tested to: IEC 60601-1:2005
EN 60601-1:2006

Certified Product: (Switching Power Supply)

License Fee - Units

Type Designation	: RPx-75y (MW)	10
x stands for	: S, D, T	1
y stands for	: 1) -3.3, -5, -12, -15, -24, -36 or -48 (x=S) 2) A or B (x=D) 3) A, B, C, D or 03 (x=T)	1
Rated Voltage	: AC 100-240V, 50/60Hz	
Rated Current	: 1.8-1.1A	
Output Voltage	: see constructional dataform	
Output Current	: see constructional dataform	
max. Ambient Temperature	: 40°C	
Protection Class	: I	


Remark: The risk management requirements of the standard were not addressed.

This certificate is valid for maximum 5 years.

Appendix: 1

12

Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signature

Dipl. Ing. (FH) M. Lohmann
TÜV Rheinland LGA Products GmbH
TÜVRheinland
Zertifizierungsstelle

Date of Issue

(day/mo/yr)

13/01/2012

Certificate



TÜV Rheinland®

This page was cancelled due to factory changed.

Please check the last page for description.

Certificate no.

TA 50220730 02

License Holder:

Mean Well Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

Mean Well (GUANGZHOU) Electronics
Co Ltd
2nd Floor, No. A Building,
Yuean Ind. Park, Dongpu Town,
TianHe Dist., Guangzhou
P.R. China

Test report no.: ZTW1-BHU 10034233 001

Client Reference: 8208400364/JH110811

Tested to: IEC 60601-1:2005
EN 60601-1:2006

Certified Product: (Switching Power Supply)

License Fee - Units

as page 01

Addition

Factory : see above

This certificate is valid for maximum 5 years.

Appendix: 1

Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signature



Dipl.-Ing. (FH) M. Schömann

Date of Issue
(day/mo/yr)

13/01/2012

Certificate



Certificate no.

TA 50220730 03

License Holder:

Mean Well Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

SuZhou Mean Well Technology
Co., Ltd.
No. 77, Jian-min Road,
Dong-qiao, Pan-yang Ind. Park,
Huang-dai Town, Xiang-cheng District
Suzhou, Jiangsu 215152
P.R. China

Test report no.: ZTW1-BHU 10034233 001

Client Reference: 8208400364/JH110811

Tested to: IEC 60601-1:2005
EN 60601-1:2006

Certified Product: (Switching Power Supply)

License Fee - Units

as page 01


Addition

Factory : see above

This certificate is valid for maximum 5 years.

Appendix: 1

Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signature LGA Product GmbH



Date of Issue

(day/mo/yr)

13/01/2012

Certificate



Certificate no.

TA 50220730 04

License Holder:

Mean Well Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

Mean Well Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Test report no.: ZTW1-SCL 10034233 002

Client Reference: 05082014_A11

Tested to: IEC 60601-1:2005
EN 60601-1:2006+A11

Certified Product: (Switching Power Supply)

License Fee - Units

as page 01

Change

Test Requirement : see above

This certificate is valid for maximum 5 years.

Appendix: 1


Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signature

W. Hsu
Dipl.-Ing. W. Hsu



Date of Issue

(day/mo/yr)
22/08/2014

Certificate



Certificate no.

TA 50220730 05

License Holder:

Mean Well Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

Mean Well Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Test report no.: ZTW1- 10034233 003

Client Reference: FL-20150706-A12

Tested to: IEC 60601-1:2005
EN 60601-1:2006+A11+A12

Certified Product: (Switching Power Supply)

License Fee - Units

as page 01

Change

Test Requirement : see above

This certificate is valid for maximum 5 years.

Appendix: 1

Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signature

Date of Issue

(day/mo/yr)

26/10/2015



Dipl.-Ing. (FH) A. Klinker

Certificate



Certificate no.

TA 50220730 06

License Holder:

MEAN WELL Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

MEAN WELL Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Test report no.: ZTW1- 10034233 004

Client Reference: ES-18042017-TUV/RPx-

Tested to: IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1+A12

Certified Product: (Medical Type Switching Power Supply)

License Fee - Units

as page 01

Change

Test Requirement : see above
License Holder : see above
Factory : see above
Product : see above

Remark: 1. The risk management requirements of the standard were not addressed;
2. This certificate is valid for maximum 5 years.

Appendix: 1

Licensed Test mark:



TÜV Rheinland Taiwan Ltd.



Signature

Dipl.-Ing. E. Stelzel

Date of Issue

(day/mo/yr)
02/02/2018

Certificate



Certificate no.

TA 50220730 07

License Holder:

MEAN WELL Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

SuZhou MEAN WELL Technology
Co., Ltd.
No. 77, Jian-min Road, Dong-qiao, Pan-
yang Ind. Park, Huang-dai Town, Xiang-
cheng District Suzhou, Jiangsu 215152
P. R. China

Test report no.: ZTW1- 10034233 004

Client Reference: ES-18042017-TUV/RPx-

Tested to: IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1+A12

Certified Product: (Medical Type Switching Power Supply)

License Fee - Units

as page 01

Change

Factory : see above

This certificate is valid for maximum 5 years.

Appendix: 1

Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signature



[Handwritten Signature]
Dipl.-Ing. F. Stoezel

Date of Issue

(day/mo/yr)

02/02/2018

Certificate



Certificate no.

TA 50220730 08

License Holder:

MEAN WELL Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

MEAN WELL (GUANGZHOU) ELECTRONICS
CO., LTD HUADU BRANCH
No. 11 Jingu South Road,
Huadong Town, Huadu District,
Guangzhou 510890
P. R. China

Test report no.: ZTW1- 10034233 004

Client Reference: ES-18042017-TUV/RPx-

Tested to: IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1+A12

Certified Product: (Medical Type Switching Power Supply)

License Fee - Units

as page 01

Addition

Factory : see above
max. Operational Altitude : 3000m

This certificate is valid for maximum 5 years.

Appendix: 1

Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signature



Dipl.-Ing. F. Stöelzel

Date of Issue

(day/mo/yr)

02/02/2018

Certificate



Certificate no.

TA 50220730 09

License Holder:

MEAN WELL Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu District,
New Taipei City 24891
Taiwan

Manufacturing Plant:

MEAN WELL Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu District,
New Taipei City 24891
Taiwan

Test report no.: ZTW1-RC 10034233 005

Client Reference: P00984755

Tested to: IEC 60601-1:2005+A1+A2
EN 60601-1:2006+A1+A12+A2

Certified Product: (Medical Type Switching Power Supply)

License Fee - Units

as page 01

Change

License Holder : see above

Test Requirement : see above

Addition

Factory : see above

Change

Factory : MEAN WELL Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

withdrawn

continued on page 10

Appendix: 3

Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signature

Date of Issue

(day/mo/yr)

02/06/2023

Dipl.-Ing. (FH) A. Klinker



Certificate



Certificate no.

TA 50220730 10

License Holder:

MEAN WELL Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu District,
New Taipei City 24891
Taiwan

Manufacturing Plant:

MEAN WELL Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu District,
New Taipei City 24891
Taiwan

Test report no.: ZTW1-RC 10034233 005

Client Reference: P00984755

Tested to: IEC 60601-1:2005+A1+A2
EN 60601-1:2006+A1+A12+A2

Certified Product: (Medical Type Switching Power Supply)

License Fee - Units

as page 01
Continuation

max. Operational Altitude : 3000m

Remark: 1. The risk management requirements of the standard
were not addressed.
2. This certificate is valid for maximum 5 years.

Appendix: 3

Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signature

Dipl.-Ing. (FH) A. Klinker

Date of Issue

(day/mo/yr)

02/06/2023



Certificate



Certificate no.

TA 50220730 11

License Holder:

MEAN WELL Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu District,
New Taipei City 24891
Taiwan

Manufacturing Plant:

Suzhou MEAN WELL Technology Co., Ltd.
No.269, Changping Road, Huangdai Town,
Xiangcheng District, Suzhou
Jiangsu
P.R. China

Test report no.: ZTW1-RC 10034233 005

Client Reference: P00984755

Tested to: IEC 60601-1:2005+A1+A2
EN 60601-1:2006+A1+A12+A2

Certified Product: (Medical Type Switching Power Supply)

License Fee - Units

as page 01

Addition

Factory : see above

This certificate is valid for maximum 5 years.

Appendix: 3

Licensed Test mark:

TÜV Rheinland Taiwan Ltd.

Signature

Date of Issue

(day/mo/yr)

02/06/2023



Dipl.-Ing. (FH) A. Klinker



Certificate



Certificate no.

TA 50220730 12

License Holder:

MEAN WELL Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu District,
New Taipei City 24891
Taiwan

Manufacturing Plant:

MEAN WELL (GUANGZHOU) ELECTRONICS
CO., LTD.
No. 11,
Jingu South Road, Huadu District,
510890 Guangzhou,
P.R. China

Test report no.: ZTW1-RC 10034233 005

Client Reference: P00984755

Tested to: IEC 60601-1:2005+A1+A2
EN 60601-1:2006+A1+A12+A2

Certified Product: (Medical Type Switching Power Supply)

License Fee - Units

as page 01

Addition

Factory : see above

This certificate is valid for maximum 5 years.

Appendix: 3

Licensed Test mark:

TÜV Rheinland Taiwan Ltd.

Date of Issue

Signature

(day/mo/yr)

02/06/2023



Dipl.-Ing. (FH) A. Klinker



Certificate



Certificate no.

TA 50220730 13

License Holder:

MEAN WELL Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu District,
New Taipei City 24891
Taiwan, R.O.C.

Manufacturing Plant:

MEAN WELL Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu District,
New Taipei City 24891
Taiwan, R.O.C.

Test report no.: ZTW1-KM 10034233 006

Client Reference: P02173322

Tested to: IEC 60601-1:2005+A1+A2
EN 60601-1:2006+A1+A2+A13

Certified Product: (Medical Type Switching Power Supply)

License Fee - Units

Change

Test Requirement : see above

This certificate is valid till 02 Feb. 2031.

Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signature

Jon S. C. Lin

Date of Issue

(day/mo/yr)

03/02/2026

